

Calibration laboratory for measure temperature and humidity  
 Kalibrierlaboratorium für die Messgröße Temperatur und Feuchte

Integrated in the certified QM - system DIN EN ISO 9001  
 Eingebunden im zertifizierten QM - System DIN EN ISO 9001

Chamber-No.:	58566078510010
<b>WKD</b>	
No.:	40224 58566078510010

## Calibration Certificate

Kalibrierschein

Object : **climatic chamber**  
 Gegenstand

Manufacturer: Vötsch Industrietechnik GmbH  
 Hersteller

Type : VC4034  
 Typ

Serial number : 58566078510010  
 Fabrikat / Serien Nr.

Inv-No. : -  
 Inventar-Nr.

Customer: NSI BVBA  
 Auftraggeber  
 Haakstraat 1A  
 B-3740 Bilzen

Order No.: -  
 Auftrags - Nr.

Number of pages of the certificate: 2  
 Anzahl der Seiten des Kalibrierscheines

Date of calibration: 15.02.2010  
 Datum der Kalibrierung

Dieser Kalibrierschein dokumentiert die Rückführung auf nationale Normale zur Darstellung der Einheiten in Übereinstimmung mit dem internationalen Einheitensystem (SI).

Für die Einhaltung einer angemessenen Frist zur Wiederholung der Kalibrierung ist der Benutzer verantwortlich .

This calibration certificate documents the traceability to national standards, which realize the units of measurement according to the International System of Units (SI).

The user is obliged to have the object recalibrated at appropriate intervals.

This calibration certificate may not be reproduced other than in full except with the permission of the WKD. Calibration certificates without signature and seal are not valid.

Dieser Kalibrierschein darf nur vollständig und unverändert weiterverarbeitet werden. Auszüge oder Änderungen bedürfen der Genehmigung des WKD. Kalibrierscheine ohne Unterschrift haben keine Gültigkeit.

Seal	Create on	Head of the service calibration department	Person incharge
Stempel	Erstell am	Leiter der Service Kalibrierabteilung	Bearbeiter



15.02.10

Mr. Reinelt  
 Phone: +49(06408)84-6343  
 Fax.: +49(06408)84-8043  
 E-Mail: t.reinelt@wut.com

Jan De Deyn  
 Issued on: \_\_\_\_\_  
 Signature

# Calibration Certificate

Chamber-No.:	58566078510010
<b>WKD</b>	
No.:	40224 58566078510010

## 1. Object of Calibration

The calibration device VC4034 is a test chamber for the size temperature and Humidity.  
For technical data please refer to the data sheet in the operating manual.

## 2. Calibration Procedure

Calibration was carried out by comparing the indications of the calibration object with the values "correct value" represented by the calibration devices/standards. The reference devices used are in valid relationship with normal / internation.

## 3. Place of Calibration

The calibration was accomplished locally at the customer.

## 4. Measuring Conditions

All measurements were made after a steady state condition of at least 2 hour/s.  
After that five measurements at 5 min. intervals have been registered. The below reference (correct value) is the arithmetic mean.

Adjusted regulation of the testing equipment: psychometric

## 5. Ambient Conditions

Temperature: 19 °C +/-5K                      Humidity: 45 %r.F. +/-20%r.F                      Pressure: 1002 hPa +/-20hPa

## 6. Measuring Results

Temperature/climate	Object of calibration		Reference		Deviation		Measuring uncertainty			
	Calibration size	Set points	Actual values	Correct value	Actual/correct value					
1.) Temperature:	-20	°C	-20,00	°C	-19,55	°C	-0,45	K	0,20	[K]
2.) Temperature:	40	°C	40,00	°C	40,10	°C	-0,10	K	0,20	[K]
Humidity:	90	%r.F	90,00	%r.F	89,94	%r.F	0,06	%r.F	1,50	[%r.F]

## 7. Measuring uncertainty

The value stated is the extended measuring uncertainty, which is the product of the standard measuring uncertainty multiplied by the extension factor k=2. This uncertainty was established in accordance with DKD-3. The value of the measured variable has a probability of 95 % within the assigned range of values. Long term stability is not considered.

## 8. Used normals

Temperature with WVPH 500 Sensor No.: 1-2                      6652-DKD-K-19501

## 9. Spatial arrangement

The measurement was made in                      **middel of testspace**

## 10. Other

Recommended date for re-calibration:                      **02.2011**

## 11. Remarks

## 12. Conformity declaration :

The measured deviations lie within the tolerance for Set point at temperature +/-1K and for the humidity +/-3 %r.F., measured in Test space center.

# Maintenance Certificate

Unit:	climatic chamber
Type:	VC4034
Series No.:	58566078510010
Order No.:	-
Customer:	NSI BVBA
Department:	Labo
Street:	Haakstraat 1A
Postcode, Place:	B-3740 Bilzen
Inventory No.:	-

Maintenance of the above unit has been carried out in accordance with the **VDMA-Standard Sheets Nos.: 24176 and 24186** and all necessary adjustments have been checked and corrected as necessary.

All values measured are within the tolerances specified by **Weiss Umwelttechnik GmbH**.

In the framework of maintenance excluding calibrated measuring instruments are used. These are checked regularly with reference standards of the enterprise. These reference standards are guaranteed by external re calibration and documented by means of a DKD-calibration certificate.

Remarks: None

Recommended date for next maintenance: 02.2011

Person responsible for maintenance: Jan De Deyn

Maintenance carried out on: 15.02.2010

Company:

Weiss Umwelttechnik GmbH  
 Simulationsanlagen Messtechnik  
 Service-Center  
 D-35447 Reiskirchen-Lindenstruth  
 Telefon (06408)84-0 Fax (06408)84-8718

Issued on: 15.02.10

Checked:

  
 \_\_\_\_\_  
 Signature

**Performed by** Jan De Deyn  
**Date** 15.02.10  
**Customer**  
**Company Name** NSI BVBA  
**Address** Haakstraat 1A  
 3740 Bilzen  
**Contact person** Karin Schepers

**Document N°** JDD121  
**Additional to Service Report :** JDD/9  
**Installation**  
**Description** climatic chamber  
**Type** VC4034  
**Serial N°** 58566078510010  
**Constr.Year** 2006

**Refrigeration Circuits**

#	Function	Refrigerant Type	Charged Weight	Stabilisation Pressure	Cooling		
					Air	Water	Cascade
1	cooling	R404A	2,5 kg	7,5 bar	X		
2			kg	bar			
3			kg	bar			

**Leak test performed on circuit :**  # 1  # 2  # 3  
 During Maintenance  After Repair  Leak Indication  New Equipment

**Procedures / Methods**

Electronic Probe

Type	TIF ZX-1E
Ser. N°	354

 UV-Light  
 Pressure Test

Start Pressure	20	bar
Test Time (hh : mm)	0:45	
End Pressure	20	bar

 Leak Seak Spray  
 Vacuum Test

Start Pressure		mbar
Test Time (hh : mm)		
End Pressure		mbar

 Open Flame Test

**Results**

No Leak Detected  
 Leak Detected

**Comments**

**Actions to Take**

Immediate repair possible  Immediate repair not possible  
 Leak controle after repair O.K. ?  Planned date for repair   
 Recuperation of refrigerant in storage cilinder  Yes  No

**Next Leak Control (Recommended)**

1 month  3 months  6 months  1 year  Other

**Summary of used refrigerant**

#	Type	Charged (kg)	Temporary Recovery (kg)	Recovery for disposal (kg)

**Signatures**

Customer :  Technician :

